



**Fit Test Adapter
Kit #38
For MSA Masks**



OHD Fit Test Adapter | Kit #38



Fig. 1



Fig. 2

INSTRUCTIONS:

1. Regulator must be removed from mask for fit testing.
2. Insert adapter into the front opening of the mask, and the O-rings will create an air-tight seal. (Fig. 1)
3. The silver prongs will prop open the inhalation valve to prepare the mask for testing. Do not twist the adapter on the mask or it could do damage to the gasket in the mask.
4. Connect the clear tubes to the two-connector side and the blue tube to the single-connector side (Fig. 2)
5. Perform fit test

CARE:

1. The adapters should be handled with care so as not to create damage that might interfere with testing.
2. Periodically check the metal hose connectors to be sure they are secure.
3. If the diaphragm should leak, remove 8 screws on top, carefully line up new diaphragm, screw in the screws loosely. Tighten opposite sides until all are secure.
4. Adapters may be cleaned with soapy water and soft cloth.
5. Adapters can be disinfected by spraying a disinfectant between tests or immersed in a cleaning solution and left out to dry. (The adapters allow outside air to be inhaled, but an employee's exhalation is routed through the exhalation valve nullifying the risk of cross-contamination.)