



OHD
YOUR PROTECTION OUR PRIORITY

**Fit Test Adapter
Kit #48
JSP Respirators**



OHD
YOUR PROTECTION OUR PRIORITY
WWW.OHDGLOBAL.COM | +1.205.980.0180

OHD Fit Test Adapter | Kit #48



Fig. 1



Fig. 2



Fig. 3

INSTRUCTIONS:

1. Make sure the gray O-rings are not missing from the black adapter (Fig. 1)
2. Line up the dimple on the adapter bottom to match the triangle on the mask (Fig. 2)
3. Push inward and turn clockwise while pushing until the rotation stops
4. Connect the clear tubes to the two-connector side and the blue tube to the single-connector side (Fig. 3)
5. Perform fit test

CARE:

1. The adapters should be handled with care so as not to create damage that might interfere with testing.
2. Periodically check the metal hose connectors to be sure they are secure.
3. If the diaphragm should leak on the single-port side of the adapter, remove 8 screws on top, carefully line up new diaphragm, screw in the screws loosely. Tighten opposite sides until all are secure.
4. Adapters may be cleaned with soapy water and soft cloth.
5. Adapters can be disinfected by spraying a disinfectant between tests or immersed in a cleaning solution and left out to dry. (The adapters allow outside air to be inhaled, but an employee's exhalation is routed through the exhalation valve nullifying the risk of cross-contamination.)